



GOVERNMENT OF ASSAM

STATE PROJECT MANAGEMENT UNIT, NEER NIRMAL PARIYOJANA, ASSAM

RWSSP-LIS, World Bank Project, 5th Floor, Redstone Building, G.S. Road, Dispur, Guwahati – 781006

Website: www.nnpphedassam.org

Phone No. 0361 – 2334345

Email ID: assam_worldbank@rediffmail.com

No. SPMU/PROC/Lab. Equip./198/2016-17/ 4475

Date: 8 February 2018

CLARIFICATION AND CORRIGENDUM 1

COUNTRY: India

NAME OF PROJECT: Neer Nirmal Pariyojana- Assam (Rural Water Supply and Sanitation Project for Low Income States- Assam)

Credit No.: 5345-IND

Contract Title: Supply, Installation and Testing of Equipment, Chemicals and Glassware for Water Quality Monitoring and Surveillance Laboratory in Project Districts of Assam under Neer Nirmal Pariyojana- Assam

Reference No.: IN-SPMU-ASSAM-22846-GO-RFB

Invitation for Bids (IFB) No.: SPMU/PROC/Lab. Equip./198/2016-17/4256

Date of IFB: 10 January 2018


The response to the Bidders' queries and the corrigendum for the above mentioned procurement are as follows:

S. No.	Bidders' Query	SPMU, NNP-Assam's Response/Corrigendum
1	BOQ: Please note that the BOQ floated in the tender is having entries like delivery dates and final destination; whereas it should have entries like price, quantity, taxes etc.	As indicated in the Tender Document
2	Manufacturer's Authorization: Please note that in Section-2 BDS (page 29), Manufacturer's Authorization is mentioned "NOT REQUIRED"; which is the key proof or certification stating whether a bidder is authorized to quote on behalf of the manufacturing concern.	As indicated in the Tender Document
3	Tender Specifications(Dimensions): Please be noted that tender specifications floated for items like Hot Plate, Bacteriological Incubator, Hot Air Oven etc have dimensions specified which will restrict the bidder to quote his optimum product in an effort to match the tendered dimensions; instead we request you to provide the capacity for the equipment to be tendered like for Hot Plate, the volume capacity, similarly for Incubators and Hot Air Ovens, the Chamber Capacity should be tendered removing	Capacity will conform to specifications. Dimensions are indicative.

S. No.	Bidders' Query	SPMU, NNP-Assam Response/Corrigendum
	the strict dimensions.	
4	Item Millipore Filtration Assembly floated in tender has mentioned the name of a manufacturer "MILLIPORE" which will restrict other bidders to quote. Hence we request you to remove the word "MILLIPORE"	The word "Millipore" is deleted
5	Item Real Time PCR Specifications: Please be noted that Real Time PCR is a very high cost equipment and requires some norms which will assist the bidder in quoting his optimum product; otherwise user may end up with products not meeting their requirements. Hence we request you to float specifications for this item, an example of which is attached herewith.	The item, Real Time PCR system as per specifications, in Line Item No. 37 of PART A: Project District-Jorhat under 1-List of Goods and Delivery Schedule of Section VII. Schedule of Requirements has been deleted.
6	Product Certifications: Please be noted that equipments floated in tender have no certification norms; which may result in procurement of products having no user safety norms, environmental factors which will not be the case with manufacturers and products having US-FDA, European CE certifications. Hence we request you to add these certifications for better product quality and efficiency.	Bureau of Indian Standards certification/other standard certification is required wherever applicable
7	Item Test Tube (Packet of 100)-(Item 51): Please note no capacity for this item has been mentioned, Hence we request you to provide some capacity value like 3ml, 5ml, 7ml etc.	Approx O.D. x Length in mm - 12x75
8	Item Porcelain Dish(Item 57): We request you to provide some approx size for the Medium Size and Large Size which is required to quote the best.	Indicated in Section VII. Schedule of Requirements

Bidders shall submit their proposals electronically only.

This Clarification and Corrigendum 1 shall form part of the bidding document.


 Project Director
 State Project Management Unit
 Neer Nirmal Pariyojana, Assam
 5th Floor, Redstone Building, G.S Road, Dispur, Guwahati – 781006